SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road,
Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

# CLASS – A INDIVIDUAL CERTIFICATE

**Certificate No. – 010150**

**Certificate Issue Date: - 16-04-2026**

**Test Method: ISO 4787 / Gravimetric Method**

| Product Detail | | | |
|---|---|---|---|
| Description of the item: Graduated Pipette Serological<br>Nominal Volume: 25 ML | | Serial No. :01:26:250<br>Tolerance: ±0.10 ML | |
| **STANDARD EQUIPMENTS USED FOR CALIBRATION** | | | |
| Name | Weighing Balance | Thermometer | |
| Make | Mettler Toledo | CE | |
| I.D. No. | WB-1 | DTM-A | |
| Range | 0-220 gm | -10°C TO 110° C | |
| Least Count | 0.1 mg | 0.1°C | |
| **CALIBRATION RESULT** | | | |
| Nominal Volume | Certified Volume | Temp. (In°C) | RH ( In % ) |
| 25 ml | 25.004 ml | 20°C | 42% |

**TERMS**

- The above product has been examined for compliance to DIN ISO 835 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

**Calibrated by:** Bindu

**Approved by:** isha

● The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders