# Scientific Instrument Traders

## CLASS – A INDIVIDUAL CERTIFICATE

**Certificate No. – 01082**

**Certificate Issue Date: - 16-03-2026**

**Test Method: ISO 4787 / Gravimetric Method**

| Product Detail | |
|---|---|
| **Description of the item: GRADUATED PIPETTE ( Mohr )** | **Serial No. :01:26:532** |
| **Nominal Volume: 1 ML** | **Tolerance: ±0.007 ML** |

**STANDARD EQUIPMENTS USED FOR CALIBRATION**

| Name | Weighing Balance | Thermometer |
|---|---|---|
| Make | Mettler Toledo | CE |
| I.D. No. | WB-1 | DTM-A |
| Range | 0-220gm. | -10°C TO 110° C |
| Least Count | 0.1 mg | 0.1°C |

**CALIBRATION RESULT**

| Nominal Volume | Certified Volume | Temp. (In°C) | RH ( In % ) |
|---|---|---|---|
| 1 ml | 1.000ml | 21°C | 42% |

**TERMS**

- The above product has been examined for compliance to DIN ISO 835 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

**Calibrated by:** Usha

**Approved by:** Shallu

● The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders

Web: www.gwsiglasslab.com E-mail: info@gwsiglasslab.com