SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road,
Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

# CLASS – A INDIVIDUAL CERTIFICATE

**Certificate No. – 011672**

**Certificate Issue Date: - 21-04-2026**

**Test Method: ISO 4787 / Gravimetric Method**

| Product Detail | |
|---|---|
| Description of the item: Measuring Cylinder | Serial No. : 01:26:222 |
| Nominal Volume: 25 ML | Tolerance: ±0.25 ML |

| STANDARD EQUIPMENTS USED FOR CALIBRATION | | |
|---|---|---|
| Name | Weighing Balance | Thermometer |
| Make | Mettler Toledo | CE |
| I.D. No. | WB-1 | DTM-A |
| Range | 0-220 gm | -10°C TO 110° C |
| Least Count | 0.1 mg | 0.1°C |

| CALIBRATION RESULT | | | |
|---|---|---|---|
| Nominal Volume | Certified Volume | Temp. (In°C) | RH ( In % ) |
| 25 ml | 24.992 ml | 20°C | 45% |

**TERMS**

- The above product has been examined for compliance to DIN ISO 4788 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

**Calibrated by:** Bindu

**Approved by:** isha

• The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders