SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road, Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

CLASS – A INDIVIDUAL CERTIFICATE

Certificate No. – 014161

Certificate Issue Date: - 02-06-2026

Test Method: ISO 4787 / Gravimetric Method

Product Detail

Description of the item: Volumetric Flask **Serial No. :01:26:711**
Nominal Volume: 100 ML **Tolerance: ±0.100 ML**

STANDARD EQUIPMENTS USED FOR CALIBRATION

Name	Weighing Balance	Thermometer
Make	Mettler Toledo	CE
I.D. No.	WB-1	DTM-A
Range	0-220gm.	-10°C TO 110° C
Least Count	0.1 mg	0.1°C

CALIBRATION RESULT

Nominal Volume	Certified Volume	Temp. (In°C)	RH (In %)
100 ml	100.024 ml	20°C	42%

TERMS

- The above product has been examined for compliance to DIN ISO 1042 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

Calibrated by: Bindu

Approved by: Isha

● The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders