SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road,
Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

# CLASS – A INDIVIDUAL CERTIFICATE

**Certificate No. – 015088**

**Certificate Issue Date: - 12-11-2026**

**Test Method: ISO 4787 / Gravimetric Method**

| Product Detail | |
|---|---|
| **Description of the item: Measuring Cylinder** | **Serial No. : 01:27:088** |
| **Nominal Volume: 25 ML** | **Tolerance: ±0.25 ML** |

| STANDARD EQUIPMENTS USED FOR CALIBRATION | | |
|---|---|---|
| **Name** | **Weighing Balance** | **Thermometer** |
| **Make** | **Mettler Toledo** | **CE** |
| **I.D. No.** | **WB-1** | **DTM-A** |
| **Range** | **0-220 gm** | **-10°C TO 110° C** |
| **Least Count** | **0.1 mg** | **0.1°C** |

| CALIBRATION RESULT | | | |
|---|---|---|---|
| **Nominal Volume** | **Certified Volume** | **Temp. (In°C)** | **RH ( In % )** |
| **25 ml** | **25.003 ml** | **20°C** | **45%** |

**TERMS**

- **The above product has been examined for compliance to DIN ISO 4788 Class A.**
- **The Product Was Found To Contain the Following Volume at reference Tem. 20°C**
- **Results reported are valid under lab condition at the time of calibration.**
- **All The Above Device Have Valid Traceability**

**Calibrated by:** Bindu

**Approved by:** isha

● **The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders**