Scientific Instrument Traders

CLASS – A INDIVIDUAL CERTIFICATE

Certificate No. – 02624

Certificate Issue Date: - 25-03-2026

Test Method: ISO 4787 / Gravimetric Method

Product Detail	
Description of the item: GRADUATED PIPETTE (Mohr)	**Serial No. :01:26:424**
Nominal Volume: 25 ML	**Tolerance: ±0.1 ML**

STANDARD EQUIPMENTS USED FOR CALIBRATION

Name	Weighing Balance	Thermometer
Make	Mettler Toledo	CE
I.D. No.	WB-1	DTM-A
Range	0-220gm.	-10°C TO 110° C
Least Count	0.1 mg	0.1°C

CALIBRATION RESULT

Nominal Volume	Certified Volume	Temp. (In°C)	RH (In %)
25 ml	25.039 ml	21°C	42%

TERMS

- The above product has been examined for compliance to DIN ISO 835 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

Calibrated by: Usha

Approved by: Shallu

● The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders

Web: www.gwsiglasslab.com E-mail: info@gwsiglasslab.com