SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road,
Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

CLASS – A INDIVIDUAL CERTIFICATE

Certificate No. – 03767

Certificate Issue Date: - 28-03-2026

Test Method: ISO 4787 / Gravimetric Method

Product Detail			
Description of the item: Volumetric Flask		**Serial No. :01:26:467**	
Nominal Volume: 250 ML		**Tolerance: ±0.150 ML**	
STANDARD EQUIPMENTS USED FOR CALIBRATION			
Name	**Weighing Balance**	**Thermometer**	
Make	**A & D**	**CE**	
I.D. No.	**WB-2**	**DTM-A**	
Range	**0-1620gm**	**-10°C TO 110° C**	
Least Count	**0.001 g**	**0.1°C**	
CALIBRATION RESULT			
Nominal Volume	**Certified Volume**	**Temp. (In°C)**	**RH (In %)**
250 ml	**249.984 ml**	**20°C**	**42%**

TERMS

- **The above product has been examined for compliance to DIN ISO 1042 Class A.**
- **The Product Was Found To Contain the Following Volume at reference Tem. 20°C**
- **Results reported are valid under lab condition at the time of calibration.**
- **All The Above Device Have Valid Traceability**

Calibrated by: Usha

Approved by: Shallu

● **The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders**