SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road,
Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

# CLASS – A INDIVIDUAL CERTIFICATE

**Certificate No. –04169**

**Certificate Issue Date: - 30-03-2026**

**Test Method: ISO 4787 / Gravimetric Method**

| Product Detail | |
|---|---|
| **Description of the item: Volumetric Flask** | **Serial No. :01:26:319** |
| **Nominal Volume: 500 ML** | **Tolerance: ±0.250 ML** |

| STANDARD EQUIPMENTS USED FOR CALIBRATION | | |
|---|---|---|
| **Name** | **Weighing Balance** | **Thermometer** |
| **Make** | **A & D** | **CE** |
| **I.D. No.** | **WB-2** | **DTM-A** |
| **Range** | **0-1620gm** | **-10°C TO 110° C** |
| **Least Count** | **0.001 g** | **0.1°C** |

| CALIBRATION RESULT | | | |
|---|---|---|---|
| **Nominal Volume** | **Certified Volume** | **Temp. (In°C)** | **RH ( In % )** |
| **500 ml** | **499.954 ml** | **20°C** | **42%** |

**TERMS**

- **The above product has been examined for compliance to DIN ISO 1042 Class A.**
- **The Product Was Found To Contain the Following Volume at reference Tem. 20°C**
- **Results reported are valid under lab condition at the time of calibration.**
- **All The Above Device Have Valid Traceability**

**Calibrated by:** Usha

**Approved by:** Shallu

● **The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders**