SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road, Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

# CLASS – A INDIVIDUAL CERTIFICATE

**Certificate No. – 06282**

**Certificate Issue Date: - 01-04-2026**

**Test Method: ISO 4787 / Gravimetric Method**

## Product Detail

**Description of the item: Volumetric Pipette** **Serial No. :01:26:232**
**Nominal Volume: 5 ML** **Tolerance: ±0.015 ML**

## STANDARD EQUIPMENTS USED FOR CALIBRATION

| Name | Weighing Balance | Thermometer |
|---|---|---|
| Make | Mettler Toledo | CE |
| I.D. No. | WB-1 | DTM-A |
| Range | 0-220gm | -10°C TO 110° C |
| Least Count | 0.1 mg | 0.1°C |

## CALIBRATION RESULT

| Nominal Volume | Certified Volume | Temp. (In°C) | RH ( In % ) |
|---|---|---|---|
| 5 ml | 5.006 ml | 20°C | 41% |

## TERMS

- The above product has been examined for compliance to DIN ISO 648 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

**Calibrated by:** Bindu

**Approved by:** isha

● The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders