SCIENTIFIC INSTRUMENT TRADERS
Precision • Quality • Trust
OEM of Laboratory Glassware
20, Pooja Vihar, Kardhan Road,
Ambala Cantt, India
www.gwsiglasslab.com
info@gwsiglasslab.com

CLASS – A INDIVIDUAL CERTIFICATE

Certificate No. – 06383

Certificate Issue Date: - 01-04-2026

Test Method: ISO 4787 / Gravimetric Method

Product Detail	
Description of the item: Volumetric Pipette	Serial No. :01:26:333
Nominal Volume: 5 ML	Tolerance: ±0.015 ML

STANDARD EQUIPMENTS USED FOR CALIBRATION

Name	Weighing Balance	Thermometer
Make	Mettler Toledo	CE
I.D. No.	WB-1	DTM-A
Range	0-220gm	-10°C TO 110° C
Least Count	0.1 mg	0.1°C

CALIBRATION RESULT

Nominal Volume	Certified Volume	Temp. (In°C)	RH (In %)
5 ml	4.996 ml	20°C	41%

TERMS

- The above product has been examined for compliance to DIN ISO 648 Class A.
- The Product Was Found To Contain the Following Volume at reference Tem. 20°C
- Results reported are valid under lab condition at the time of calibration.
- All The Above Device Have Valid Traceability

Calibrated by: Bindu

Approved by: isha

● The Certificate Shall Not Be Reproduced Except in Full And Without Written Permission From Scientific Instrument Traders